

AED REQUIREMENTS

- A Certificate of Medical Oversight is placed next to the AED unit.

Current expiration of Certificate of Medical Oversight is March 1, 2024

- Monthly Monitoring through a third- party documentation.

We use AED Total solutions-Shondra Scott monitors them.

AED California State requirements:

- AED Registration is required.

Julian Corral (Occupational Health Coordinator EH&S) confirmed that all four units are registered.

- AEDs should be maintained according to the manufacturer's specifications and tested at least biannually and after use.
- When an AED is placed in a building, the building owner shall, at least once a year, notify the tenants as to the location of the AED units and provide information to tenants about who they can contact if they want to take AED or CPR training.
- Instructions for AED use should be posted in 14-point type next to the device.

AED Cabinet

- AED Cabinets have an internal alarm system, this is to alert staff to respond to the area. There is a shunt key to turn off the alarm. Keys: Hernan has 1, Floor 2 has key in fire extinguisher case in their lobby, Nichole Sparks has two keys in key case, Jonathan can access.

SHCC Building has 4 AED machines.

New units-Lobby AED machines monitored monthly.

Floor 1 lobby - SN-A23H-00598-Adult pads expiration 11/30/2025-Lot-C053023-05

Floor 2 lobby - SN-A23H-00205-Adult pads expiration 11/30/2025-Lot-C053023-05

Older Clinic AED units, monitored monthly.

Station 1-SN-A15H-03686-Adult pads expiration 07/03/2025-Lot-CO10323-02

Trauma room-SN-A16F-06546-Adult pads expiration 6/27/2025-Lot- C122722-02